



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to
Complete this report whenever the instrument is serviced or repaired and whenever it is plac
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

RECEIVED
By Carol Day at 3:35 pm, Nov 26, 2013

DATAMASTER SN 940234	NAME OF AGENCY Crocker Police Department	DATE OF INSPECTION 11/21/2013
LOCATION OF INSTRUMENT (STREET AND CITY) 108 South Commercial Crocker MO 65452		TIME OF INSPECTION 9:49 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 11/21/2013 21:53

☒ COMPUTER ☒ DETECTOR

☒ PROGRAM ☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 48.°C ☒ QUARTZ STANDARD

☒ FLOW DETECTOR ☒ CALIBRATION

☒ PUMP HIGH SPEED ☒ PRINTER

☒ INDICATOR LIGHTS

☒ SIMULATOR SOLUTION SUPPLIER Rep Co Marketing Inc LOT # 12002 EXP. DATE 08/29/2014

☒ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2729 EXP. DATE 08/28/2014

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099

TEST 2 .099

TEST 3 .099

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 3	(.05-.09) 1	(.10-.14) 1	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE

PRINT FULL NAME
Aaron J Baker

TYPE II PERMIT NUMBER/EXPIRATION DATE
230112 06/12/2015

TELEPHONE NUMBER
(573) 736-2211

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd,
Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-189 STONYBROOK DRIVE
RALEIGH, N.C. 27604
819-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain 1209 gms/dl \pm .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 \pm 3% gms/210L Breath when heated to 34 Degrees Celsius \pm 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CROCKER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 142204
11/21/13

ARREST TIME: 21:13
SUBJECT NAME:
REFTEST
DOB: 11/11/11 SEX: M
STATE/C.L.: NA/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: NA
TESTING OFFICER:
BAKER/ARON/J
OFFICER I.D.: 506
PERMIT NUMBER: 203112
EXPIRATION DATE: 06/12/15
MISCELLANEOUS DATA:
REFTEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:16
INTERNAL STANDARD	VERIFIED	20:16
RADIO INTERFERENCE		

Operator Signature



506



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

AARON J BAKER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
577.020 through 577.041, RSMo and 306.131 through 306.119, RSMo.

DATE 06/12/2013

NUMBER 230112

EXPIRES 06/12/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

W. J. Baker
Shad V. Vukobratovic

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAS-4 (Rev-10)

MO S&D-0771 (8-10)